

TAMIL NADU AGRICULTURAL UNIVERSITY

Office of the Director
Directorate of Centre for Plant Protection Studies
TNAU, Coimbatore – 641 003
Date: 18.08.2016

No. TNAU/ CPPS/ AEN/ NADP /Refrigerated centrifuge / Purchase/Corrigendum/ 2016

Tender for purchase of refrigerated centrifuge – Pre bid meeting

Venue : Directorate of Centre for Plant Protection Studies
TNAU, Coimbatore – 641 003

Date and Time : 18.08.2016 at 04.00 PM

Scheme : GOI – NADP(RKVY) Scheme

Consequent upon the discussions held during the Pre-bid meeting the following points are considered for incorporation with suitable amendment

Corrigendum

Particulars	As per tender notice	Now amended
Max RCF	20,000 X g	20,000 X g and above
Rotor locking system	Install and remove rotor without tool in less than 5 seconds with just a push of a button, for quick and easy change of rotors for different applications	Install and remove rotor without tool in less than 5 seconds with just a push of a button, for quick and easy change of rotors for different applications or other safer locking systems
Primary Rotor	Fixed angle rotor for spinning 15 ml and 50 ml tubes(Manufactured from a highly corrosion and fatigue resistant carbon fiber material warranted for at least 12 years)	(Fixed angle rotor for spinning 6 – 14 x 15 ml and 50 ml tubes)Manufactured from a highly corrosion and fatigue resistant carbon fiber / other metals
Additional Rotors	Fixed angle rotor for spinning 30 x 1.5 ml and 2.0 ml tubes	Fixed angle rotor for spinning 24 -30 x 1.5 ml and 2.0 ml tubes
Type of Motor/Drive Power	Asynchronous motor (Brushless) Low power consumption - 200 w	Asynchronous motor (Brushless) Low power consumption

Antony 18.08.16
Professor and Head
Department of Agricultural Entomology
PROFESSOR AND HEAD
Dept. of Agrl. Entomology
CPPS, TNAU, CBE-3

K. Ramani 18/8/16
Director
Centre for Plant Protection Studies
DIRECTOR
Centre for Plant Protection Studies
Tamil Nadu Agricultural University
COIMBATORE - 641 003